

FeLV-FIV Diagnostic Kit

Uranotest

For veterinary use only

Technical basis

The URANOTEST FeLV-FIV diagnostic kit is based on the immunochromatographic technique and is designed for the qualitative detection of *Feline Immunodeficiency virus* (FIV) and *Feline Leukaemia virus* (FeLV) in feline whole blood, serum and plasma.

The test has a double structure; it contains two single tests: a strip for FeLV antigen detection and a strip for FIV antibody detection. Each test consists of several overlapped membranes. On one of the membranes, there are a test line (T line) and control line (C line). The lines are not visible before applying the sample. After applying the sample in the appropriate sample well, migration begins by capillarity action through the membrane. If the result is negative, one purple colour band appears in the C area. This line, called control line, always appears, as it is a control line indicating that the test has successfully performed. If the test result is positive, in addition to the control line, a second line will form in the test area (Test line).

Materials supplied

- 1 - Double test devices individually packaged in aluminium pouch.
- 2 - Dropper bottle with buffer solution.
- 3 - Disposable capillary pipettes for sample collection. The mark at the capillary indicates a volume of 10 µl.



- 4 - Vials with anticoagulant (EDTA) for blood collection.
- 5 - Instructions for use.

Precautions

- 1 - For veterinary use only.
- 2 - Wear disposable gloves when handling the samples. All samples should be treated as potentially infectious. Wash and disinfect hands after handling. Avoid aerosol formation when dispensing the sample.
- 3 - To obtain good results, it is important to add the correct sample volume.
- 4 - Open the device just before use.
- 5 - All reagents must be at room temperature before performing the test.
- 6 - Do not use the test if the envelope is damaged or broken.
- 7 - Do not re-use.
- 8 - Do not use reagents after the expiry date.
- 9 - Do not use kits with different batch numbers.
- 10 - The quality of each component of the kit has been individually assessed for each batch. Do not mix components or reagents from kits with different batch numbers.

Preservation and stability

The kit must be stored at a temperature between 2 and 30°C. Under these conditions, we can guarantee the stability until the expiry date printed on the box and on the individual pouch.

The kit has been developed to be stored at room temperature. Although it also can be stored in the refrigerator, we recommend store it at room temperature to avoid the need to wait for reagents to reach the room temperature.

DO NOT FREEZE. Do not exposure to direct sunlight.

Sample collection and preparation

The test can be performed with serum, plasma or whole blood (treated with anticoagulant).

WHOLE BLOOD

Take a sample of blood using traditional clinical methods in a tube containing anticoagulant (heparin, EDTA or citrate). The kit includes EDTA tubes; however, any of the aforementioned anticoagulants can be used.

The blood should be analysed within 4 hours after extraction. If is not possible, it can be kept cold between 2 and 8 °C for no more than 24 hours. Do not freeze.

Haemolysed samples may affect the results.

PLASMA

Take a sample of blood using traditional clinical methods in a tube containing anticoagulant (heparin, EDTA or citrate).

Separate the plasma by centrifugation. The plasma can be kept refrigerated at a temperature between 2 and 8°C up to 72 hours. For conservation over a longer period, it should be frozen under -20°C. If the sample has been refrigerated, wait for it to reach room temperature before testing.

SERUM

Take a sample of blood using traditional clinical methods in a tube without anticoagulant.

Separate the serum by centrifugation. The serum can be kept refrigerated at a temperature between 2 and 8°C up to 72 hours. For conservation over a longer period, it should be frozen under -20°C. If the sample has been refrigerated, wait for it to reach room temperature before testing.

Instructions for use

- 1 - Remove the test device from the protective pouch and place it on a flat and dry surface.
- 2 - Using the capillary pipette provided, transfer 10 µl of sample (whole blood, serum, or plasma) into the round FeLV sample well. Add 2 drops of buffer solution on the FeLV sample well.
- 3 - Repeat exactly the same procedure for the FIV test.


- 4 - When the test begins running, you will observe migration of sample through the result window. If migration has not begun after 1 minute, add one more drop of buffer solution.
- 5 - Read the results within 5-10 minutes. Do not read results after 20 minutes. Coloured lines appeared after 20 minutes have not diagnostic value and should be ignored.

FIV determination zone

10 µl of sample



2 drops of
buffer solution



Interpret at 10 minutes



FeLV determination zone

10 µl of sample



2 drops of
buffer solution



Interpret at 10 minutes



Interpreting results


1 - Negative result

There is only a single purple line on the C area, both in the FIV test and in the FeLV test. This line should always appear.




4 - Positive result for Leukaemia

Two lines appear on the result window of the FeLV test (T and C lines), but only a single line appears on the result window of FIV test (C line).




2 - Positive result for Immune Deficiency and Leukaemia simultaneously

Two lines appear on both result windows. The test is positive to FeLV when two lines appear on the result window (T and C line). In the same way, the test is positive to FIV when two lines appear on the result window of the FIV test device. Whichever line appears first, the result is considered positive.






5 - Invalid result

The test is invalid if not coloured line appears at the Control area (C) even if a coloured line appears in the Test area (T). The reason may be due to incorrect handling or using a damaged test.



3 - Positive Result for Immune Deficiency

Two lines appear on the result window of the FIV test (T and C lines), but only a single line appears on the result window of FeLV test (C line).



Limitations of the technique

Even though the URANOTEST FeLV-FIV diagnostic kit shows high sensitivity and specificity, cannot be excluded a low incidence of false positive or negative results.

As any other laboratory procedure, the definitive clinical diagnosis cannot be based only on the test result. It must be based on an ensemble of clinical and laboratory procedures. If there is any doubt, repeat the test and/or contrast with other diagnostic methods.